



Date of Birth:
 Device: **AUTOGEN X4 CRT-D G179**
 Clinic:
 Search Tags:
 Tachy Mode: **Monitor + Therapy**

Latest Device Transmission: **22 Sep 2022 11:38 CEST**
 Last Office Interrogation: **23 May 2022**
 nt Date:
 Patient Group:

V-3040: 22 May 2022 11:32, VT, A Rate: 385 min⁻¹, V Rate: 168 min⁻¹

Detail

VT Event Onset

Avg A Rate 385 min⁻¹
 Avg V Rate 168 min⁻¹
 Detection Onset/Stability
 Onset Percent
 Template N/R

At Inhibit

Stability 34 ms
 V>A Rate False
 AFib True
 RhythmMatch™ N/R

Attempt 1, Burst V ATP

Elapsed Time 00:00:47
 Inhibited Attempt
 ATP Information
 Number of Bursts 3

At V-Detect

Avg A Rate 368 min⁻¹
 Avg V Rate 180 min⁻¹
 Rate Zone VT
 RhythmMatch™ N/R
 SRD Met (False, Off)
 ATP Timeout False
 Onset Intvl (0 ms, Off)
 Onset % 0 %

Attempt 2, Ramp V ATP

Elapsed Time 00:01:17
 ATP Information
 Number of Bursts 3

At V-Detect



Avg A Rate 296 min⁻¹
 Avg V Rate 187 min⁻¹
 Rate Zone VT
 ATP Timeout False


Attempt 3, 41 J V Shock


Elapsed Time 00:01:46
 Shock Information
 Charge Time 11.9 s
 Lead Impedance 57 Ω
 Lead Polarity Initial

Event Ended 00:02:30



EGM displayed at 25mm per second










AFibV Unstb AFibV
V-Dur Unstb
Unstb AFibV




















↑ > V Range	↓ < V Range	■ V Sense	□ V Pace
↑ > A Range	↓ < A Range	• A Sense	○ A Pace

