



Date of Birth:
 Device: **MOMENTUM X4 CRT-D**
 Clinic:
 Search Tags:
 Tachy Mode: **Monitor + Therapy**

Latest Device Transmission: **23 Jan 2023 05:01 CET**
 Last Office Interrogation: **19 Oct 2019**
 Implant Date:
 Patient Group:

V-1: 18 Oct 2019 17:00, VF, A Rate: 172 min⁻¹, V Rate: 259 min⁻¹

Detail

VF Event Onset

Avg A Rate	172 min ⁻¹
Avg V Rate	259 min ⁻¹
Detection	Rate Only
Template	N/R

At V-Detect

Avg A Rate	108 min ⁻¹
Avg V Rate	147 min ⁻¹
Rate Zone	VF
Stability	(113 ms, Off)
V>A Rate	(True, Off)
AFib	(True, Off)
RhythmMatch™	N/R
SRD Met	(False, Off)
ATP Timeout	False
Onset Intvl	(0 ms, Off)
Onset %	(0 %, Off)

Attempt 1, Non Therapy V

Elapsed Time 00:00:15

Attempt Information

No therapy delivered, Tachy therapy mode is not programmed to Monitor+Therapy.

Event Ended

00:00:33

EGM displayed at 25mm per second











