






V-2: 02 Jul 2022 19:26, VF, A Rate: 267 min⁻¹, V Rate: 278 min⁻¹



Detail


VF Event Onset			
Avg A Rate	267 min ⁻¹		
Avg V Rate	278 min ⁻¹		
Detection	Onset/Stability		
Onset	Percent		
Template	N/R		
At V-Detect			
Avg A Rate	283 min ⁻¹	Attempt 1, 41 J V Shock	
Avg V Rate	280 min ⁻¹	Elapsed Time	00:00:03
Rate Zone	VF	Aborted Attempt	
Stability	(37 ms, Off)	No shock attempted due to fail to reconfirm	
V>A Rate	(False, Off)		
AFib	(True, Off)		
RhythmMatch™	N/R		
SRD Met	(False, Off)		
ATP Timeout	False		
Onset Intvl	(374 ms, Off)		
Onset %	61 %		
At V-Detect			
Avg A Rate	283 min ⁻¹	Attempt 2, 41 J V Shock	
Avg V Rate	226 min ⁻¹	Elapsed Time	00:00:11
Rate Zone	VF	Shock Information	
ATP Timeout	False	Charge Time	6.2 s
		Lead Impedance	64 Ω
		Lead Polarity	Reversed
Event Ended			00:00:49
EGM displayed at 25mm per second			











↑ > V Range	↓ < V Range	■ V Sense	□ V Pace
↑ > A Range	↓ < A Range	• A Sense	○ A Pace

