




V-54: 12 Nov 2018 16:40, VF, A Rate: 98 min⁻¹, V Rate: 235 min⁻¹


Detail

VF Event Onset			
Avg A Rate	98 min ⁻¹		
Avg V Rate	235 min ⁻¹		
Detection	Rhythm ID		
Template	12 Nov 2018 14:47		
RhythmMatch™ Threshold	94 %		
At V-Detect			
Avg A Rate	77 min ⁻¹	Attempt 1, 41 J V Shock	
Avg V Rate	229 min ⁻¹	Elapsed Time	00:00:03
Rate Zone	VF	VF ATP delivered prior to shock	
Stability	(5 ms, Off)	Shock Information	
V>A Rate	(True, Off)	Charge Time	7.3 s
AFib	(False, Off)	Lead Impedance	61 Ω
RhythmID Correlated	False	Lead Polarity	Initial
RhythmMatch™	Too Fast		
SRD Met	(False, Off)		
ATP Timeout	False		
At V-Detect			
Avg A Rate	108 min ⁻¹	Attempt 2, 41 J V Shock	
Avg V Rate	231 min ⁻¹	Elapsed Time	00:00:36
Rate Zone	VF	Shock Information	
Stability	(8 ms, Off)	Charge Time	7.0 s
V>A Rate	(True, Off)	Lead Impedance	62 Ω
AFib	(False, Off)	Lead Polarity	Initial
SRD Met	(False, Off)		
ATP Timeout	False		
Event Ended			00:01:14
EGM displayed at 25mm per second			

















↑ > V Range	↓ < V Range	■ V Sense	□ V Pace
↑ > A Range	↓ < A Range	• A Sense	○ A Pace

