



ZOOM® View™ DISK

Selected Episodes Report

Detection 3,

Date of Birth

Device

Tachy Mode

TELIGEN 100 F110/456789

Monitor + Therapy

Report Created 01 Jun 2011

Last Office Interrogation

6 Sep 2011

Implant Date

1 Jan 2009

Event V-13386: 24 Aug 2011 07:22

VF Event Onset

Avg A Rate: 56 bpm

Avg V Rate: 321 bpm

Detection: Rhythm ID

At V-Detect

Avg A Rate: 54 bpm

Avg V Rate: 111 bpm

Rate Zone: VF

Stability: (258 ms, Off)

V>A Rate: (True, Off)

AFib: (False, Off)

RhythmID Correlated: False

SRD Met: (False, Off)

ATP Timeout: False

Attempt 1, 41 J V Shock

Elapsed Time: 00:00:03

Aborted Attempt

VF ATP delivered, no shock attempted due to fail to reconfirm

At V-Detect

Avg A Rate: 58 bpm

Avg V Rate: 199 bpm

Rate Zone: VT

ATP Timeout: False

Attempt 2, Burst V ATP

Elapsed Time: 00:00:15

ATP Information:

Number of Bursts: 1

Event Ended 00:00:28











